



INLAND VACUUM INDUSTRIES, INC.
35 Howard Avenue
Churchville, NY 14428
www.InlandVacuum.com

Certificate of Analysis

Inland 19

Lot Number 13356
Manufacture Date 12/10/25

Characteristic	Specification	Result
Viscosity @ 40 °C	49.0 - 61.0 cSt	56.2 cSt
Density	0.865 - 0.890	0.87
Ultimate pressure (mechanical)	$\leq 1.0 \times 10^{-3}$ torr	pass
Color	<1.0	pass

Notice: All data given herein are accurate and presented without warranty or responsibility of any kind expressed or implied. The customer shall not hold Inland Vacuum Industries Inc. liable for any problems resulting from the use and handling of this material.

Authorization _____

Date _12/10/25